Vol. 02, Iss. 2 (2024) 125-135, DOI: 10.61552/JME.2024.02.004

JME

Journal of Materials and Engineering

www.jme.aspur.rs

Review article

Recent Advances in Bio-Tribology From Joint Lubrication to Medical Implants: A Review

Shalom Akhai[a], Amandeep Singh Wadhwa[b,*]

[a]*Chandigarh College of Engineering, Jhanjeri-140307, India,*
[b]*UIET, Panjab University, Chandigarh-160014, India.*

Keywords:

Bio-tribology
Joint Lubrication
Medical Implants
Wear Analysis
Surface Engineering
Biomimetics
Computational Modeling

* Corresponding author:

Amandeep Singh Wadhwa
E-mail:
aman_wadhwa77@rediffmail.com

Received: 18 November 2023
Revised: 24 December 2023
Accepted: 12 January 2024

A B S T R A C T

Bio-tribology, the study of lubrication, wear, and friction in biological systems, has made recent strides in the lubrication of medical implants and joints. This comprehensive review discusses the most recent developments in bio-tribology, with an emphasis on the transition from basic science to medical engineering applications. An overview of bio-tribology and its increasing significance in medicine commences the study. Following this, recent research on the composition of synovial fluid and its function in minimizing friction in natural joints is discussed. Medical implant tribology, including degradation issues in dental prostheses and joint replacements, is then addressed. This study investigates the manner in which surface modifications and coatings improve the tribological performance of medical implants. It emphasizes recent developments in material science and engineering. This article emphasizes the significance of minimizing implant material degradation in relation to the functionality and longevity of biomedical devices. Following this, digital models, simulations, and state-of-the-art imaging techniques that have propelled bio-tribology forward are highlighted. In addition to proposing potential research avenues to overcome the current obstacles, the article highlights the interdisciplinary nature of bio-tribology and urges collaboration among tribology, materials science, and biomechanics researchers.

© 2024 Journal of Materials and Engineering

1. INTRODUCTION

The multidisciplinary science of bio-tribology investigates the mechanisms underlying lubrication, wear, and friction in biological systems [1,2]. The investigation of the intricate mechanisms that support natural processes is an essential area of research that has far-reaching consequences for medical applications [3,4]. As an offshoot of the tribological and biological sciences, bio-tribology investigates the fundamental mechanisms that regulate friction and wear at biological interfaces, including joints [5-7]. This entails the analysis of complex biological tribosystems in order to decipher the underlying molecular and biomechanical principles. The potential of bio-tribology to

fundamentally transform the design and functionality of biomedical devices is its significance in medical applications [8,9]. A comprehension of the operation of natural joints has served as a catalyst for the creation of biomimetic medical implant solutions. The utilisation of bio-tribological principles is particularly significant when it comes to medical implants, as the reduction of wear, enhancement of lubrication, and guarantee of biocompatibility are of the utmost importance [10-14].This review provides a detailed analysis of recent advancements in bio-tribology, focusing on medical implant tribology and joint lubrication. It synthesizes data from various research initiatives to offer a comprehensive outlook on the field's current state. The review aims to provide valuable insights for engineers, medical professionals, and researchers, enhancing their understanding of bio-tribology and its potential for technology and medicine.

2. JOINT LUBRICATION

The following section delves into the investigation of joint lubrication, an essential domain within bio-tribology, with an emphasis on its intricacies, the composition of synovial fluid, and the advancements achieved in imitating natural joint lubrication for novel medical purposes. It emphasises the continuous progress that is being made in this particular field.

2.1 Recent Developments in Understanding Tribological Aspects

Recent progress in the comprehension of tribological elements pertaining to natural joint lubrication has yielded valuable knowledge regarding the intricate relationship between biological constituents situated within joints [15]. The utilisation of magnetic resonance imaging and high-resolution microscopy has enabled scientists to observe and analyse the intricate biomechanics associated with joint lubrication. This has resulted in a more comprehensive comprehension of the mechanisms by which natural joints attain their wear-resistant and low-friction properties [16,17]. Biomechanical investigations have additionally examined the impact of physiological conditions, muscle activity, and joint morphology on tribological performance, thereby offering a comprehensive outlook on the complexities of joint lubrication from a tribological viewpoint [18,19].

2.2 Synovial Fluid Composition and Its Role in Friction Reduction

As an essential element in lubricating joints, synovial fluid has been the focus of biotribological investigations. Recent research has unveiled the multifunctional characteristics of this substance, including rheological properties that are specifically designed to accommodate the demands of joint movement [20-22]. Understanding the molecular composition of synovial fluid constituents, such as lubricin and hyaluronic acid, has yielded valuable information regarding the prevention of wear and friction [23]. An intriguing subject of investigation is the dynamic characteristics of synovial fluid, which include its ability to adapt to changing joint conditions and react to mechanical stimuli [24].Compositional knowledge of synovial fluid aids in the comprehension of natural joint lubrication and the development of synthetic lubricants intended for medical use.

2.3 Advancements in Mimicking Natural Joint Lubrication

The imitation of natural joint lubrication for medical applications is the focus of research. The use of biomimicry in the development of synthetic joints and lubricants is a prominent area of research. The synthesis of lubricants that emulate the composition and rheological properties of synovial fluid has demonstrated potential in mitigating wear and friction in artificial joints. The objective of surface modifications and coatings applied to medical implants is to emulate the tribological properties exhibited by natural joint tissues [25-27]. By incorporating biomimetic materials, such as hydrogels that possess adjustable lubricating properties, joint implant longevity and performance could be improved. Technological advances in 3D printing permit the construction of intricate structures.

In general, this section highlights the evolving nature of joint lubrication research, which has evolved from understanding the tribological complexities of natural joints to developing innovative medical interventions, emphasizing the potential of bio-tribological insights to revolutionize the field of medical joint lubrication [28,29].

3. MEDICAL IMPLANT TRIBOLOGY

Medical implant tribology covers wear, friction, and lubrication in joint replacements and dental prosthesis. It examines surface modifications, coatings, and case studies that demonstrate creative materials and designs to enhance these implants' tribological performance, emphasising materials science, biomechanics, and bio-tribology [30].

3.1 Tribological Challenges

The effectiveness and endurance of medical implants, including dental prosthetics and joint replacements, are significantly impacted by tribological challenges. The difficulties encompass reducing wear, reducing friction, and assuring the enduring stability of interfaces between the host and implant [31,32]. In addition to physiological variations and complex loading conditions, prospective adverse tissue reactions can have an effect on the tribological performance of implants. Joint replacements encounter difficulties pertaining to the interface between materials and tissues, whereas dental implants confront distinct tribological challenges stemming from the corrosive conditions of the oral cavity and masticatory forces [33-35]. Comprehending and effectively confronting these obstacles is critical in enhancing the longevity and performance of implants.

3.2 Surface Modifications and Coatings

In order to enhance the tribological performance of medical implants, a multitude of surface modifications and coatings are being investigated by researchers and engineers. Orthopaedic implants feature biocompatible coatings, such as hydroxyapatite, which promote osseointegration and decrease degradation. Coatings composed of nanocomposites reduce wear and friction in dental implants [36]. In joint replacements, diamond-like carbon (DLC) coatings provide enhanced resistance to wear and decreased friction. By incorporating lubricating materials and self-lubricating coatings, artificial lubrication mechanisms in joints are replicated. These developments facilitate the creation of medical implants that possess improved tribological properties in addition to withstanding mechanical stresses [37].

4. WEAR AND WEAR PARTICLE ANALYSIS

The analysis of wear in bio-tribological systems is pivotal for understanding the performance and longevity of medical implants. This section scrutinizes recent methodologies employed in wear analysis, explores the implications of wear debris, and highlights studies focused on minimizing wear in implant materials.

4.1 Recent Methods for Analyzing Wear in Bio-Tribological Systems

The field of wear analysis has been significantly transformed by developments in analytical techniques, which now permit exhaustive and accurate examinations of wear in bio-tribological systems [38,39]. Surface profilometry, scanning electron microscopy, atomic force microscopy, and wear simulation models are methodologies that provide valuable insights into the mechanisms of wear, alterations in surface topography, and progressive material degradation [40-42]. Methods for in-situ wear testing and wear particle characterization enable the observation of tribological processes occurring at bio-tribological interfaces in real time. The utilisation of microscale and nanoscale imaging technologies in wear studies facilitates the comprehension of intricate interplays that occur between biological tissues and implant materials [43].

4.2 Implications of Wear Debris in the Context of Medical Implants

The presence of debris from wear within bio-tribological systems has the potential to greatly affect the functionality and compatibility of medical implants. Osteolysis, inflammatory reactions, and possible implant loosening are all potential consequences [44]. It is of the utmost importance to comprehend the dimensions, makeup, and biological responsiveness of wear particles in order to forecast the enduring impacts on adjacent tissues and the operation of the implant. Recent research has investigated the relationship between wear debris and the immune system of the host, uncovering inflammatory pathways in the process [45,46]. Gaining insight into the biological reactions induced by wear can facilitate the formulation of approaches to alleviate the detrimental consequences and guarantee the effectiveness and safety of implants.

4.3 Focusing on Minimizing Wear in Implant Materials

Innovative material science, surface engineering, and design solutions to decrease implant material wear are examined. Zirconia and alumina provide hardness and wear resistance, while nanocomposites and biomimetic coatings improve tribology [47,48]. Research on biological and synthetic lubrication, biomimetic and lubricin analogue border lubricants, implant geometries, and articulating surfaces may minimise wear, improve performance, and extend implant lifetime.

Table 1. Wear-Resistant Coatings and Articulating Surfaces for Hip and Knee Implants.

Article Title	Year	Reference	Summary
Current Status and Future Potential of Wear-Resistant Coatings and Articulating Surfaces for Hip and Knee Implants	2022	[49]	Provides a comprehensive overview of wear-resistant coatings for joint replacements, including various coatings that have been used or are under investigation to minimize wear in implant materials.
Mobile-Bearing Total Knee Arthroplasty: Design Factors in Minimizing Wear	2006	[50]	Discusses design factors that can minimize wear in total knee arthroplasty (TKA), emphasizing the potential benefits of mobile-bearing TKA, which can reduce long-term polyethylene wear.
The Influence of Material and Design on Total Knee Replacement Wear	2011	[51]	Examines the influence of implant design on wear reduction in total knee replacement (TKR), comparing the wear resistance of different knee systems under physiologic stair-climb loading and motion profiles.
Retrieval Analysis of TiN (Titanium Nitride) Coated Knee Replacements: Coating Wear and Degradation in Vivo	2020	[52]	Examines the in vivo wear and degradation of titanium nitride (TiN) coatings used in knee replacements, finding evidence of coating wear and roughening on tibial bearing surfaces.
The Lexicon of Polyethylene Wear in Artificial Joints	2007	[53]	Provides an overview of the vocabulary used to describe wear in polyethylene components of artificial joints, discussing the different wear modes, mechanisms, damage types, and debris generated in prosthetic joints.

5. EMERGING TECHNOLOGIES IN BIO-TRIBOLOGY

Technology has advanced bio-tribology research by giving new techniques to study biological friction, wear, and lubrication. Optimising tribological performance requires recent bio-tribology advances including computer models, simulations, and improved imaging.

Table 2. Emerging Trends in Biotribology - From Biosensors to Biodegradable Metals.

Advancements in Bio-Tribology Research	Description	Reference	Year	Applications
Development of bioinspired surfaces for reducing friction and wear	Scientists are studying the tribological behavior of bioinspired surfaces, such as those found on lotus leaves and shark skin, to develop new materials with improved wear resistance and biocompatibility.	[54]	2021	Medical implants, prosthetics, artificial joints, biomedical devices
Investigation of sliding wear properties of sustainable biocomposites	Sustainable biocomposites are being investigated for their potential as biomaterials with improved sliding wear properties.	[55]	2023	Medical implants, prosthetics, artificial joints, biomedical devices, sustainable biomaterials
Use of active agents and biomaterials to improve bio-lubrication and strengthen soft tissues	Active agents, such as enzymes and nanoparticles, are being used to improve the lubrication of soft tissues, while biomaterials are being used to strengthen soft tissues and improve wear resistance.	[56]	2023	Soft tissue engineering, wound healing, cartilage repair, osteoarthritis treatment

Utilization of graphene derivatives and nanocomposites in tribology and lubrication	Graphene derivatives and nanocomposites are being used to develop new biolubricants and biomaterials with improved tribological properties.	[57]	2019	Medical implants, prosthetics, artificial joints, biomedical devices, biolubricants
Additive manufacturing of biodegradable metals for biomedical applications	Additive manufacturing is being used to create porous biodegradable metals with improved biocompatibility and tribological properties.	[58]	2019	Bone implants, tissue engineering scaffolds, drug delivery devices, biomedical implants
Application of biosensors in tissue engineering	Biosensors are being used to monitor the biotribological performance of engineered tissues, allowing for real-time feedback on the effectiveness of the tissue and the need for adjustments.	[59]	2014	Tissue engineering, medical implants, drug delivery

5.1 Recent Technological Advancements in Bio-Tribology Research

Progress in bio-tribology technology has broadened the scope of investigation and enhanced our comprehension of the tribological mechanisms at work in biological systems [60]. Sophisticated methods for surface characterization, such as high-resolution microscopy and spectroscopy, facilitate in-depth examinations of tribological interfaces at the micro and nanoscale. Technological and material science advancements have enabled the creation of innovative biomaterials that possess customised tribological properties [61]. The incorporation of intelligent materials enables biological systems to dynamically adapt to shifting tribological conditions [62].

5.2 Role of Computational Models in Optimizing Tribological Performance

Computational models and simulations play a pivotal role in the field of bio-tribology research as they enable the examination of intricate interactions across multiple dimensions [63,64]. These models facilitate the comprehension of the tribological characteristics of biological interfaces by researchers, including the atomic-level interactions that dictate friction and wear in biomolecules. Macroscopic simulations offer valuable insights into the mechanical behaviour of joints and implants when subjected to various loading conditions [65-68]. Multibody dynamics simulations and finite element analysis aid in the comprehension of stress distribution and contact mechanics in complex biological systems. The design of biomimetic surfaces and materials is guided by the prediction and optimisation of tribological performance made possible by the integration of computational approaches [69].

5.3 Imaging Techniques in Understanding and Optimizing Tribological Performance

Sight-capturing real-time tribological events in biological systems requires the utilisation of sophisticated imaging techniques, such as scanning probe microscopy, confocal microscopy, and high-speed imaging [70]. By enabling the observation of dynamic processes, such as the formation of fluid films in joints, high-speed imaging provides insight into lubrication mechanisms. By enabling three-dimensional imaging of biological tissues and tribological interfaces, confocal microscopy enables a comprehensive examination of surface characteristics and wear patterns [71]. Scanning probe microscopy methods, such as scanning tunnelling microscopy (STM) and atomic force microscopy (AFM), provide the capability to analyse surface topography and mechanical properties at the nanoscale. These technologies enhance the comprehension of tribological phenomena occurring in biological systems and facilitate the optimisation of synthetic tribological interfaces utilised in biomechanical and medical applications [72,73].

6. CHALLENGES AND FUTURE DIRECTIONS IN BIO-TRIBOLOGY

Though bio-tribology has made progress, this section analyses its limitations, offers future study, and considers its connection with materials science and biomechanics.

6.1 Current Challenges in Bio-Tribology

The tribological complexity of living systems presents various obstacles for bio-tribology. Biomedical interfaces are dynamic, making it hard to recreate and comprehend complicated joint interactions, which delays clinical translation [74]. Variable bio-tribological systems have variable mechanical and physiological circumstances, requiring a context-specific methodology. Wear debris and host immunological response are additional issues. Experimental procedures and measuring methods must be standardised for repeatability and comparability [75]. Bridging these gaps is crucial for moving the science forward and applying discoveries to medical implants.

6.2 Future Research and Development Avenues

Addressing present issues and developing new research pathways will advance bio-tribology. Understanding the molecular and biomechanical details of natural tribological systems, such as proteins and signalling networks, may help explain the basics. Biomaterials with tailorable characteristics and smart materials that respond to physiological situations may improve artificial joints and implants. Advanced computer models and AI can forecast and optimise tribological behaviour, revealing critical wear, friction, and lubrication aspects. Nanotechnology and nano-tribology enable molecular and nanoscale manipulation and control of tribological interactions, opening new bio-tribological study avenues.

6.3 Integration with Other Disciplines

Bio-tribology must be integrated with materials science and biomechanics for comprehensive research and development [76]. Tribologists, materials scientists, and biomechanical engineers may advance together. Materials science develops biomaterials and surface changes, whereas biomechanics studies joint and implant function. Experimental results may be better interpreted when biomechanical context is understood. Cross-disciplinary research may improve medical implant design and performance by using bio-tribological information [77]. Bio-tribology advances by addressing difficulties, discovering new research pathways, and developing multidisciplinary partnerships.

7. CONCLUSION

This article analysed the evolving environment of bio-tribology, including challenges and developments. Natural joint lubrication, medical implant tribology, wear analysis, and emerging technologies were all subjects that were further explored. Surface modifications, coatings, and biomimetic designs enhance the biocompatibility and resistance to wear of medical implants. For the analysis of wear and wear particles, sophisticated imaging and modelling tools provide unique insights into the tribological behaviour of bio-tribological systems. It is critical to comprehend the impact of wear debris on medical implants in order to mitigate adverse effects and prolong the lifespan of the devices. Computational models and improved imaging have significantly transformed the field of bio-tribology research by offering novel perspectives on the tribological performance of biological systems. Bio-tribology challenges, such as the complexity of biological interfaces and wear debris, require continued research. Understanding molecular interactions, developing intelligent biomaterials, and applying computer models to further the field should be the next steps. The design, development, and optimisation of medical implants are guided by bio-tribology in order to maximise their durability and functionality while minimising wear and friction concerns. Bio-tribological insights must be incorporated into medical implant technologies in order to enhance patient outcomes, minimise revision operations, and elevate quality of life.

REFERENCES

[1] J. F. Kayode, S. I. Lawal, S. A. Afolalu, "Overview of green tribology in recent world: fundamentals and future development," in *2023 International Conference on Science, Engineering and Business for Sustainable Development Goals*, vol. 1, pp. 1-11, IEEE, 2023, [Online]. Available: doi: 10.1109/SEB-SDG57117.2023.10124520.

[2] K. Srivastava, J. S. Rathore, S. Shrivastava, "Multidimensional outline of experimental techniques for human skin tribology: a scoping review," *Proceedings of the Institution of Mechanical Engineers, Part C: Journal of Mechanical Engineering Science*, p. 09544062231209824, 2023.

[3] F. Di Puccio and L. Mattei, "Biotribology of artificial hip joints," *World Journal of Orthopedics*, vol. 6, no. 1, pp. 77-94, 2015, doi: 10.5312/wjo.v6.i1.77.

[4] M. J. Nine, D. Choudhury, A. C. Hee, R. Mootanah, and N. A. A. Osman, "Wear debris characterization and corresponding biological response: Artificial hip and knee joints," *Materials*, vol. 7, no. 2, pp. 980-1016, 2014, doi: 10.3390/ma7020980.

[5] R. M. F. Wagner, R. Maiti, M. J. Carré, C. M. Perrault, P. C. Evans, R. Lewis, "Bio-tribology of vascular devices: a review of tissue/device friction research," *Biotribology*, vol. 25, p. 100169, 2021, [Online]. Available: doi: 10.1016/j.biotri.2021.100169.

[6] M. Marian, D. Berman, D. Nečas, N. Emami, A. Ruggiero, A. Rosenkranz, "Roadmap for 2D materials in biotribological/biomedical applications–a review," *Advances in Colloid and Interface Science*, vol. 307, pp. 102747, 2022.

[7] J. Sun, S. Du, "Application of graphene derivatives and their nanocomposites in tribology and lubrication: a review," *RSC Advances*, vol. 9, no. 69, pp. 40642-40661, 2019, [Online]. Available: doi: 10.1039/C9RA05679C.

[8] Z. R. Zhou, Z. M. Jin, "Biotribology: Recent progresses and future perspectives," *BioSurface and Biotribology*, vol. 1, no. 1, pp. 3-24, 2015, [Online]. Available: doi: 10.1016/j.bsbt.2015.03.001.

[9] S. Jahn, J. Seror, and J. Klein, "Lubrication of articular cartilage," *Annual Review of Biomedical Engineering*, vol. 18, pp. 235–258, 2016, doi: 10.1146/annurev-bioeng-081514-123305.

[10] J. K. Hirwani, S. K. Sinha, "Bio-tribological studies of Structalit/UHMWPE composites as an alternative to UHMWPE for hip joint application," *Wear*, vol. 518, pp. 204630, 2023, [Online]. Available: doi: 10.1016/j.wear.2023.204630.

[11] S. Bhoi, A. Prasad, A. Kumar, R. B. Sarkar, B. Mahto, C. S. Meena, C. Pandey, "Experimental study to evaluate the wear performance of UHMWPE and XLPE material for orthopedics application," *Bioengineering*, vol. 9, no. 11, pp. 676, 2022, [Online]. Available: doi: 10.3390/bioengineering9110676.

[12] T. Pratap, K. Patra, "Micro–nano surface texturing, characterization, and their impact on biointerfaces," in *Advanced Machining and Finishing*, pp. 577-610, Elsevier, 2021.

[13] R. M. Ronchi, H. G. De Lemos, R. K. Nishihora, M. G. D. V. Cuppari, S. F. Santos, "Tribology of polymer-based nanocomposites reinforced with 2D materials," *Materials Today Communications*, pp. 105397, 2023, [Online]. Available: doi: 10.1016/j.mtcomm.2023.105397.

[14] M. Q. Chen, "Recent advances perspective of nanotechnology-based implants for orthopedic applications," *Frontiers in Bioengineering and Biotechnology*, vol. 10, pp. 878257, 2022, [Online]. Available: doi: 10.3389/fbioe.2022.878257.

[15] R. Rufaqua, M. Vrbka, D. Choudhury, D. Hemzal, I. Křupka, M. Hartl, "A systematic review on the correlation between biochemical and mechanical processes of lubricant film formation in joint replacement of the last 10 years," *Lubrication Science*, vol. 31, no. 3, pp. 85-101, 2019.

[16] L. Eck, M. Yang, J. J. Elias, C. S. Winalski, F. Altahawi, N. Subhas, X. Li, "Quantitative MRI for evaluation of musculoskeletal disease: cartilage and muscle composition, joint inflammation, and biomechanics in osteoarthritis," *Investigative Radiology*, vol. 58, no. 1, pp. 60-75, 2023, [Online]. Available: doi: 10.1097/RLI.0000000000000909.

[17] M. Tschaikowsky, S. Brander, V. Barth, R. Thomann, B. Rolauffs, B. N. Balzer, T. Hugel, "The articular cartilage surface is impaired by a loss of thick collagen fibers and formation of type I collagen in early osteoarthritis," *Acta Biomaterialia*, vol. 146, pp. 274-283, 2022, [Online]. Available: doi: 10.1016/j.actbio.2022.04.036.

[18] K. Moghadasi, M. S. M. Isa, M. A. Ariffin, S. Raja, B. Wu, M. Yamani, M. R. B. Muhamad, F. Yusof, M. F. Jamaludin, M. S. bin Ab Karim, N. bin Yusoff, "A review on biomedical implant materials and the effect of friction stir-based techniques on their mechanical and tribological properties," *Journal of Materials Research and Technology*, vol. 17, pp. 1054-1121, 2022, [Online]. Available: doi: 10.1016/j.jmrt.2022.01.050.

[19] A. Siddaiah, P. L. Menezes, "Advances in bio-inspired tribology for engineering applications," *Journal of Bio- and Tribo-Corrosion*, vol. 2, pp. 1-19, 2016, [Online]. Available: doi: 10.1007/s40735-016-0053-0.

[20] C. Pradal, G. E. Yakubov, M. A. Williams, M. A. McGuckin, J. R. Stokes, "Lubrication by biomacromolecules: mechanisms and biomimetic strategies," *Bioinspiration & Biomimetics*, vol. 14, no. 5, p. 051001, 2019, [Online]. Available: doi: 10.1088/1748-3190/ab2ac6.

[21] G. W. Greene, D. W. Lee, J. Yu, S. Das, X. Banquy, J. N. Israelachvili, "Lubrication and wear protection of natural (bio) systems," in *Polymer Adhesion, Friction, and Lubrication*, pp. 83-133, 2013, [Online]. Available: doi: 10.1002/9781118505175.ch3.

[22] R. Shah, B. Gashi, S. Hoque, M. Marian, A. Rosenkranz, "Enhancing mechanical and biomedical properties of prostheses-surface and material design," *Surfaces and Interfaces*, vol. 27, p. 101498, 2021, [Online]. Available: doi: 10.1016/j.surfin.2021.101498.

[23] W. Lin, J. Klein, "Recent progress in cartilage lubrication," *Advanced Materials*, vol. 33, no. 18, p. 2005513, 2021, [Online]. Available: doi: 10.1002/adma.202005513.

[24] T. Alliston, C. J. Hernandez, D. M. Findlay, D. T. Felson, O. D. Kennedy, "Bone marrow lesions in osteoarthritis: what lies beneath," *Journal of Orthopaedic Research*, vol. 36, no. 7, pp. 1818-1825, 2018, [Online]. Available: doi: 10.1002/jor.23844.

[25] S. Kaur, K. Ghadirinejad, R. H. Oskouei, "An overview of the tribological performance of titanium alloys with surface modifications for biomedical applications," *Lubricants*, vol. 7, no. 8, p. 65, 2019, [Online]. Available: doi: 10.3390/lubricants7080065.

[26] S. Ghosh, S. Abanteriba, "Status of surface modification techniques for artificial hip implants," *Science and Technology of Advanced Materials*, vol. 17, no. 1, pp. 715-735, 2016, [Online]. Available: doi: 10.1080/14686996.2016.1240575.

[27] Z. A. Uwais, M. A. Hussein, M. A. Samad, N. Al-Aqeeli, "Surface modification of metallic biomaterials for better tribological properties: A review," *Arabian Journal for Science and Engineering*, 42, pp. 4493-4512, 2017, [Online]. Available: doi: 10.1007/s13369-017-2624-x.

[28] A. Ruggiero, "Milestones in natural lubrication of synovial joints," *Frontiers in Mechanical Engineering*, vol. 6, p. 52, 2020, [Online]. Available: doi: 10.3389/fmech.2020.00052.

[29] D. Necas, M. Vrbka, M. Marian, B. Rothammer, S. Tremmel, S. Wartzack, A. Galandakova, J. Gallo, M. A. Wimmer, I. Křupka, M. Hartl, "Towards the understanding of lubrication mechanisms in total knee replacements–Part I: Experimental investigations," *Tribology International*, vol. 156, p. 106874, 2021, [Online]. Available: doi: 10.1016/j.triboint.2021.106874.

[30] P. Sahoo, S. K. Das, J. P. Davim, "Tribology of materials for biomedical applications," in *Mechanical Behaviour of Biomaterials*, pp. 1-45, Woodhead Publishing, 2019.

[31] S. Affatato, F. Traina, "Bio and medical tribology," in *Tribology for Engineers*, Woodhead Publishing, pp. 243-286, 2011.

[32] D. Bitar, J. Parvizi, "Biological response to prosthetic debris," *World Journal of Orthopedics*, vol. 6, no. 2, p. 172, 2015, [Online]. Available: doi: 10.5312/wjo.v6.i2.172.

[33] F. Eltit, Q. Wang, R. Wang, "Mechanisms of adverse local tissue reactions to hip implants," *Frontiers in Bioengineering and Biotechnology*, vol. 7, p. 176, 2019, [Online]. Available: doi: 10.3389/fbioe.2019.00176.

[34] S. Elleuch, H. Jrad, M. Wali, F. Dammak, "Finite element analysis of the effect of porosity on the biomechanical behaviour of functionally graded dental implant," *Proceedings of the Institution of Mechanical Engineers Part E: Journal of Process Mechanical Engineering*, p. 09544089231197857, 2023, [Online]. Available: doi: 10.1177/09544089231197857.

[35] S. Affatato, M. Spinelli, M. Zavalloni, C. Mazzega-Fabbro, M. Viceconti, "Tribology and total hip joint replacement: current concepts in mechanical simulation," *Medical Engineering & Physics*, vol. 30, no. 10, pp. 1305-1317, 2008, [Online]. Available: doi: 10.1016/j.medengphy.2008.07.006.

[36] S. M. Baligidad, S. Abdullah, D. K. Rajendran, G. Thodda, "Investigation on fatigue life of HAp and rGO hybrid coating on the PEEK artificial screws for dental implant," *Materials Chemistry and Physics*, vol. 309, p. 128381, 2023, [Online]. Available: doi: 10.1016/j.matchemphys.2023.128381.

[37] M. Z. Ibrahim, A. A. Sarhan, F. Yusuf, M. Hamdi, "Biomedical materials and techniques to improve the tribological, mechanical and biomedical properties of orthopedic implants–A review," *Journal of Alloys and Compounds*, vol. 714, pp. 636-667, 2017, [Online]. Available: doi: 10.1016/j.jallcom.2017.04.231.

[38] J. Jamari, M. I. Ammarullah, A. P. M. Saad, A. Syahrom, M. Uddin, E. Van der Heide, H. Basri, "The effect of bottom profile dimples on the femoral head on wear in metal-on-metal total hip arthroplasty," *Journal of Functional Biomaterials*, vol. 12, no. 2, p. 38, 2021, [Online]. Available: doi: 10.3390/jfb12020038.

[39] H. A. Abdel-Aal, "Functional surfaces for tribological applications: inspiration and design," *Surface Topography: Metrology and Properties*, vol. 4, no. 4, p. 043001, 2016, [Online]. Available: doi: 10.1088/2051-672X/4/4/043001.

[40] R. M. Trommer, M. M. Maru, W. L. Oliveira Filho, V. P. S. Nykanen, C. P. Gouvea, B. S. Archanjo, E. M. Ferreira, R. F. Silva, C. A. Achete, "Multi-scale evaluation of wear in UHMWPE-metal hip implants tested in a hip joint simulator," *Biotribology*, vol. 4, pp. 1-11, 2015.

[41] Hoque, M. A., Yao, C. W., Khanal, M., *Lian, I.*, "Tribocorrosion Behavior of Micro/Nanoscale Surface Coatings," *Sensors*, vol. 22, no. 24, p. 9974, 2022. [Online]. Available: doi: 10.3390/s22249974.

[42] M. S. Uddin, "Wear measurement and assessment of explanted cross-linked PE acetabular cups using a CMM," *Tribology Transactions*, vol. 57, no. 5, pp. 767-777, 2014, doi: 10.1080/10402004.2014.911398.

[43] Zimmermann, E. A., Fiedler, I. A., Busse, B., "Breaking new ground in mineralized tissue: Assessing tissue quality in clinical and laboratory studies," *Journal of the Mechanical Behavior of Biomedical Materials*, vol. 113, p. 104138, 2021. [Online]. Available: doi: 10.1016/j.jmbbm.2020.104138.

[44] Goodman, S. B., Gallo, J., Gibon, E., Takagi, M., "Diagnosis and management of implant debris-associated inflammation," *Expert review of medical devices*, vol. 17, no. 1, pp. 41-56, 2020.

[45] Connors, J. P., Stelzer, J. W., Garvin, P. M., Wellington, I. J., Solovyova, O., "The Role of the Innate Immune System in Wear Debris-Induced Inflammatory Peri-Implant Osteolysis in Total Joint Arthroplasty," *Bioengineering*, vol. 9, no. 12, p. 764, 2022. [Online]. Available: doi: 10.3390/bioengineering9120764.

[46] Guder, C., Gravius, S., Burger, C., Wirtz, D. C., Schildberg, F. A., "Osteoimmunology: A current update of the interplay between bone and the immune system," *Frontiers in Immunology*, vol. 11, p. 58, 2020. [Online]. Available: doi: 10.3389/fimmu.2020.00058.

[47] Farooq, S. A., Raina, A., Mohan, S., Arvind Singh, R., Jayalakshmi, S., Irfan Ul Haq, M., "Nanostructured coatings: Review on processing techniques, corrosion behaviour and tribological performance," *Nanomaterials*, vol. 12, no. 8, p. 1323, 2022. [Online]. Available: doi: 10.3390/nano12081323.

[48] Salari, M., Mohseni Taromsari, S., Bagheri, R., Faghihi Sani, M. A., "Improved wear, mechanical, and biological behavior of UHMWPE-HAp-zirconia hybrid nanocomposites with a prospective application in total hip joint replacement," *Journal of Materials Science*, vol. 54, no. 5, pp. 4259-4276, 2019. [Online]. Available: doi: 10.1007/s10853-018-3146-y.

[49] Skjoldebrand, C., Tipper, J. L., Hatto, P., Bryant, M., Hall, R. M., Persson, C., "Current status and future potential of wear-resistant coatings and articulating surfaces for hip and knee implants," *Materials Today Bio*, vol. 15, p. 100270, 2022. [Online]. Available: doi: 10.1016/j.mtbio.2022.100270.

[50] Dennis, D. A., Komistek, R. D., "Mobile-bearing total knee arthroplasty: design factors in minimizing wear," *Clinical Orthopaedics and Related Research*, vol. 452, pp. 70-77, 2006. [Online]. Available: doi: 10.1097/01.blo.0000238776.27316.d6.

[51] Essner, A., Herrera, L., Hughes, P., Kester, M., "The influence of material and design on total knee replacement wear," *The journal of knee surgery*, pp. 009-018, 2011. [Online]. Available: doi: 10.1055/s-0031-1275390.

[52] Lapaj, L., Rozwalka, J., "Retrieval analysis of TiN (titanium nitride) coated knee replacements: Coating wear and degradation in vivo," *Journal of Biomedical Materials Research Part B: Applied Biomaterials*, vol. 108, no. 4, pp. 1251-1261, 2020. [Online]. Available: doi: 10.1002/jbm.b.34473.

[53] McKellop, H. A., "The lexicon of polyethylene wear in artificial joints," *Biomaterials*, vol. 28, no. 34, pp. 5049-5057, 2007. [Online]. Available: doi: 10.1016/j.biomaterials.2007.07.040.

[54] Hoskins, J. K., Zou, M., "Designing a Bioinspired Surface for Improved Wear Resistance and Friction Reduction," *Journal of Tribology*, vol. 143, no. 5, p. 051107, 2021. [Online]. Available: doi: 10.1115/1.4050673.

[55] H. Pulikkalparambil, A. Babu, A. Thilak, N. P. Vighnesh, S. M. Rangappa, S. Siengchin, "A review on sliding wear properties of sustainable biocomposites: Classifications, fabrication and discussions," *Heliyon*, 2023. [Online]. Available: doi: 10.1016/j.heliyon.2023.e14381.

[56] T. Garg, G. Sharma, S. Shankar, S. P. Singh, "Tribological performance of polymeric materials for biomedical applications," *Assessment of Polymeric Materials for Biomedical Applications*, CRC Press, pp. 121-138, 2023.

[57] A. Çolak, M. Göktaş, and F. Mindivan, "Effect of reduced graphene oxide amount on the tribological properties of UHMWPE biocomposites under water-lubricated conditions," *SN Applied Sciences*, vol. 2, no. 3, Feb. 2020, doi: 10.1007/s42452-020-2179-4.

[58] Y. Qin, P. Wen, H. Guo, D. Xia, Y. Zheng, L. Jauer, R. Poprawe, M. Voshage, J. H. Schleifenbaum, "Additive manufacturing of biodegradable metals: Current research status and future perspectives," *Acta biomaterialia*, vol. 98, pp. 3-22, 2019. [Online]. Available: doi: 10.1016/j.actbio.2019.04.046.

[59] A. Hasan, M. Nurunnabi, M. Morshed, A. Paul, A. Polini, T. Kuila, M. Al Hariri, Y. K. Lee, A. A. Jaffa, "Recent advances in application of biosensors in tissue engineering," *BioMed Research International*, 2014. [Online]. Available: doi: 10.1155/2014/307519.

[60] P. Sahoo, S. K. Das, "Tribology—a tool for mechanical and industrial engineering," *Mechanical and Industrial Engineering: Historical Aspects and Future Directions*, pp. 1-37, 2022. [Online]. Available: doi: 10.1007/978-3-030-90487-6_1.

[61] S. Kumar, M. Nehra, D. Kedia, N. Dilbaghi, K. Tankeshwar, K. H. Kim, "Nanotechnology-based biomaterials for orthopaedic applications: Recent advances and future prospects," *Materials Science and Engineering: C*, vol. 106, p. 110154, 2020. [Online]. Available: doi: 10.1016/j.msec.2019.110154.

[62] H. Gong, C. Yu, L. Zhang, G. Xie, D. Guo, J. Luo, "Intelligent lubricating materials: A Review," *Composites Part B: Engineering*, vol. 202, p. 108450, 2020. [Online]. Available: doi: 10.1016/j.compositesb.2020.108450.

[63] P. M. Johns-Rahnejat, R. Rahmani, H. Rahnejat, "Current and future trends in tribological research," *Lubricants*, vol. 11, no. 9, p. 391, 2023. [Online]. Available: doi: 10.3390/lubricants11090391.

[64] A. Kurdi, N. Alhazmi, H. Alhazmi, T. Tabbakh, "Practice of simulation and life cycle assessment in tribology—A review," *Materials*, vol. 13, no. 16, p. 3489, 2020. [Online]. Available: doi: 10.3390/ma13163489.

[65] Y. Li, C. Ma, "A multiscale computational framework for wear prediction in knee replacement implants," *Mechanics of Materials*, vol. 175, p. 104480, 2022. [Online]. Available: doi: 10.1016/j.mechmat.2022.104480.

[66] Y. Heriveaux, S. Le Cann, M. Fraulob, E. Vennat, V. H. Nguyen, G. Haïat, "Mechanical micro modeling of stress-shielding at the bone-implant interphase under shear loading," *Medical & Biological Engineering & Computing*, vol. 60, no. 11, pp. 3281-3293, 2022. [Online]. Available: doi: 10.1007/s11517-022-02657-2.

[67] Z.-H. Zhang, Y.-S. Qi, B.-G. Wei, H.-R.-C. Bao, and Y.-S. Xu, "Application strategy of finite element analysis in artificial knee arthroplasty," *Frontiers in Bioengineering and Biotechnology*, vol. 11, p. 1127289, 2023, doi: 10.3389/fbioe.2023.1127289.

[68] L. Gao, X. Lu, X. Zhang, Q. Meng, and Z. Jin, "Lubrication modelling of artificial joint replacements: Current status and future challenges," *Lubricants*, vol. 10, no. 10, p. 238, 2022, doi: 10.3390/lubricants10100238.

[69] S. Badini, S. Regondi, R. Pugliese, "Unleashing the power of artificial intelligence in materials design," *Materials*, vol. 16, no. 17, p. 5927, 2023. [Online]. Available: doi: 10.3390/ma16175927.

[70] L. Wang, Y. Yang, Z. Liu, J. Tian, Y. Meng, T. Qi, T. He, D. Li, P. Yan, M. Gong, Q. Liu, "High-speed all-fiber micro-imaging with large depth of field," *Laser & Photonics Reviews*, vol. 16, no. 9, p. 2100724, 2022. [Online]. Available: doi: 10.1002/lpor.202100724.

[71] J. M. Rosas, D. J. Atkins, A. L. Chau, Y. T. Chen, R. Bae, M. K. Cavanaugh, R. I. Espinosa Lima, A. Bordeos, M. G. Bryant, A. A. Pitenis, "In vitro models of soft tissue damage by implant-associated frictional shear stresses," *Proceedings of the Institution of Mechanical Engineers, Part J: Journal of Engineering Tribology*, vol. 237, no. 5, pp. 1264-1271, 2023. [Online]. Available: doi: 10.1177/13506501221132897.

[72] K. Gulati and T. Adachi, “Profiling to probing: Atomic force microscopy to characterize nano-engineered implants,” *Acta Biomaterialia*, vol. 170, pp. 15–38, 2023, doi: 10.1016/j.actbio.2023.08.006.

[73] I. S. Bayer, "Advances in tribology of lubricin and lubricin-like synthetic polymer nanostructures," *Lubricants*, vol. 6, no. 2, p. 30, 2018. [Online]. Available: doi: 10.3390/lubricants6020030.

[74] K. Prasad, O. Bazaka, M. Chua, M. Rochford, L. Fedrick, J. Spoor, R. Symes, M. Tieppo, C. Collins, A. Cao, D. Markwell, "Metallic biomaterials: Current challenges and opportunities," *Materials*, vol. 10, no. 8, p. 884, 2017. [Online]. Available: doi: 10.3390/ma10080884.

[75] W. Al-Zyoud, D. Haddadin, S. A. Hasan, H. Jaradat, O. Kanoun, "Biocompatibility Testing for Implants: A Novel Tool for Selection and Characterization," *Materials*, vol. 16, no. 21, p. 6881, 2023. [Online]. Available: doi: 10.3390/ma16216881.

[76] J. D. B. De Mello, H. L. Costa, "Brazilian Tribology: origin, status quo and future perspectives," *Surface Topography: Metrology and Properties*, vol. 11, no. 3, p. 030201, 2023. [Online]. Available: doi: 10.1088/2051-672X/ace83e.

[77] A. Lubbeke, J. A. Smith, D. Prieto-Alhambra, A. J. Carr, "The case for an academic discipline of medical device science," *EFFORT Open Reviews*, vol. 6, no. 3, pp. 160-163, 2021. [Online]. Available: doi: 10.1302/2058-5241.6.200094.